

1 **B-50 Replace Battery Banks, \$71,700**

2 Q. Provide a copy of the most recent reports resulting from routine maintenance
3 tests and inspections of the battery bank to be replaced at each of the Stony
4 Brook and Western Avalon Terminal Station.

5

6

7 A. Attached are copies of the preventive maintenance inspection reports for the
8 Stony Brook and Western Avalon Terminal station batteries for the period
9 2000 to the present.

Stoney Brook Battery PM'S

2006 PM'S

Equipment ID: _____
Manufacturer: G N BStation: StoneySerial No.: 90A-09Date: 2006-01-17 Readings By: JS+BSRoom Temperature: 18 °CPosition Found: ☐ float ☐ equalizePosition Left: ☐ float ☐ equalizeVoltage of Battery Disconnected: 128.8 VPilot Cell Voltage: 2.2 VPilot Cell No.: 218Pilot Cell Specific Gravity: N/APilot Cell Temp.: 18°Water Added: ☐ yes ☒ noEqualize Charge: ☐ yes ☒ no

(specific gravity should be maintained between 1200 & 1220)

Voltage Output of Charger (V): 130 Found130 LeftDC Current Output of Charger (A): 6 Found6 LeftElectrolyte Level: ☐ low ☐ very low ☐ normalCabinet Heater: ☐ on ☐ off

CHECKS COMPLETED:

- ☒ Check the accuracy of the charger voltage meter against another meter.
- ☒ Check each cell for cracks, corrosion on terminals, leaks.
- ☒ Inspect the battery rack for signs of corrosion & battery leakage.
- ☐ Check cells for deposits on the bottom
- ☐ Check the operation of ventilation fans and thermostat.
- ☒ Check the operation of emergency lighting.

Note Any Concerns Below:

Equalizing Charge			
Why given:			
Voltage:	Amps:	Duration:	hrs

**Notes:

125V Bank: Float - 129V Equalize - 139V

48V Bank: Float - 51.6V Equalize - 56V

-Equalize charge is required after any water is added

-Do not allow excessive gassing

System ID: _____
 Manufacturer: GAB

Station: Stone 4
 Date: 2006-01-17

Serial No.: 90A-09
 Readings By: TS+BS

128.8 VDC

	Volt	Temp (°C)	Specific Gravity	Electrolyte Level			Volt.	Temp (°C)	Specific Gravity	Electrolyte Level	
				(low / normal / high)	Refilled					(low / normal / high)	Refilled
1	2.2	18°C				31	2.2				
2	2.2					32	2.2				
3	2.2					33	2.2				
4	2.2					34	2.2				
5	2.2					35	2.2				
6	2.2					36	2.2				
7	2.2					37	2.2				
8	2.2					38	2.2				
9	2.2					39	2.2				
10	2.2					40	2.2				
11	2.2					41	2.2				
12	2.2					42	2.2				
13	2.2					43	2.2				
14	2.2					44	2.2				
15	2.2					45	2.2				
16	2.2					46	2.2				
17	2.2					47	2.2				
18	2.2					48	2.2				
19	2.2					49	2.2				
20	2.2					50	2.2				
21	2.2					51	2.2				
22	2.1					52	2.2				
23	2.2					53	2.2				
24	2.2					54	2.2				
25	2.2					55	2.2				
26	2.2					56	2.2				
27	2.2					57	2.2				
28	2.2					58					
29	2.2					59					
30	2.2					60					

** Take temperature readings at the negative terminal of cells.

Take specific gravity readings before adding water to cells.

Add water to cells with low levels of electrolyte.

2005 PM'S

Equipment ID: BATTERYS
Manufacturer: GNBStation: STB
Date: 05-03-28Serial No.: _____
Readings By: GN/WS

Room Temperature: _____ °C

Position Found: ☒ float ☐ equalize
Position Left: ☒ float ☐ equalizeVoltage of Battery Disconnected: — VPilot Cell Voltage: 2.4 V
Pilot Cell Specific Gravity: N/APilot Cell No.: _____
Pilot Cell Temp.: — N/AWater Added: ☐ yes ☐ no DRY
Equalize Charge: ☐ yes ☐ no
(specific gravity should be maintained between 1200 & 1220)Voltage Output of Charger (V): 130 Found 130 Left
DC Current Output of Charger (A): 6 Found 6 LeftElectrolyte Level: ☐ low ☐ very low ☐ normal
Cabinet Heater: ☐ on ☐ off N/A

CHECKS COMPLETED:

- OK Check the accuracy of the charger voltage meter against another meter.
- OK Check each cell for cracks, corrosion on terminals, leaks.
- OK Inspect the battery rack for signs of corrosion & battery leakage.
- N/A Check cells for deposits on the bottom
- OK Check the operation of ventilation fans and thermostat.
- OK Check the operation of emergency lighting.

Note Any Concerns Below:

Equalizing Charge			
Why given:			
Voltage:	Amps:	Duration:	hrs

**Notes:

125V Bank: Float - 129V Equalize - 139V
 48V Bank: Float - 51.6V Equalize - 56V

- Equalize charge is required after any water is added
- Do not allow excessive gassing

microCELLTRON DataLogge

Notes:

STB

2005-03-28

23/28/05 10:27

Reference Value Entered:
#--- Siemens (MHOS)Reference Temperature:
12 Degrees Celsius

----String Statistics----

String Average : 867s
String% of HI Jar: 073%

-----Jar Statistics-----

JAR	JAR--V	JAR--G	+STRAP
001	2.24V	0829s	0887s
002	2.20V	1060s	0797s
003	2.24V	0829s	1093s
004	2.25V	1102s	0884s
005	2.25V	0842s	0866s
006	2.24V	0931s	0607s
007	2.24V	0614s	0786s
008	2.24V	0764s	1008s
009	2.25V	0882s	0937s
010	2.25V	0904s	0852s
011	2.25V	0889s	0827s
012	2.25V	0896s	0661s
013	2.25V	0665s	0967s
014	2.25V	1028s	0712s
015	2.27V	0715s	0754s
016	2.27V	0639s	0887s
017	2.27V	0835s	0813s
018	2.25V	0835s	1025s
019	2.22V	0967s	0973s
020	2.24V	0963s	1062s
021	2.25V	1082s	0826s
022	2.25V	0829s	0862s
023	2.25V	0878s	0884s
024	2.24V	0914s	0950s
025	2.25V	0881s	0557s
026	2.24V	0620s	0818s
027	2.24V	0835s	0828s
028	2.25V	0852s	0895s
029	2.24V	0849s	0892s
030	2.25V	0846s	0829s
031	2.25V	0914s	1023s
032	2.25V	0860s	1035s
033	2.25V	1118s	0664s
034	2.24V	0692s	0856s
035	2.25V	0971s	1056s
036	2.24V	1018s	0919s
037	2.25V	0859s	0877s
038	2.24V	0943s	0942s
039	2.25V	1050s	1022s
040	2.24V	0963s	0626s
041	2.22V	0620s	1035s
042	2.25V	0922s	0584s
043	2.22V	0610s	0810s
044	2.25V	0835s	0868s
045	2.25V	0931s	0804s
046	2.25V	0790s	0670s
047	2.24V	0720s	0846s
048	2.22V	0846s	0594s
049	2.22V	0608s	0962s
050	2.25V	0986s	0803s
051	2.24V	0875s	0578s
052	2.22V	0634s	0937s
053	2.25V	0970s	1046s
054	2.25V	1086s	0862s
055	2.24V	0889s	0919s
056	2.25V	1031s	0473s
057	2.22V	0729s	0000s

Lowest Jar:
049Highest Jar:
004

---Setup Information---

String Warn% (?): 70%
String Fail% (?): 60%
Jar Warn% (?): 70%
Jar Fail% (?): 60%
Jar Selected: 02V

2004 PM'S

~~RECORDED~~ 710 No. 1 Brook

MANUEL IVK HDSOLYTE

CAT. NO. 90A-09

TYPE ELECTROLYTE Dry Sulfuric Acid

CHARGING CONDITION FOUND:

FLOAT ☒ EQUALIZE ☐ VOLTS 128.9 AMPS 6

PILOT CELL:

NO. 78 (FOUND) SP. GRAVITY MA VOLTS 2.22 TEMP — DEG C

WATER ADDED: YES ☐ NO ☒ EQUALIZE CHARGE: VOLTS — AMP — HR. —
 BATH CLEANED yes CLEAN AND INTERDATE CONNECTIONS yes BATH TESTED yes
 NEW PILOT CELL: NO. — SP. GRAVITY —
 CHECK ACCURACY OF THERMOMETERS: yes OK

RECORD ALL CELLS:

SP GR	VOLT	TEMP	SP GR	VOLT	TEMP	SP GR	VOLT	TEMP
1.		<u>15c</u>	11.		<u>15c</u>	21.		<u>15c</u>
2.			12.			22.		
3.			13.			23.		
4.			14.			24.		
5.			15.			25.		
6.			16.			26.		
7.			17.			27.		
8.			18.			28.		
9.			19.			29.		
10.			20.			30.		
11.			21.			31.		
12.			22.			32.		
13.			23.			33.		
14.			24.			34.		
15.			25.			35.		
16.			26.			36.		
17.			27.			37.		
18.			28.			38.		
19.			29.			39.		
20.			40.			40.		

ONLY IF INSTRUCTED:

TERMINATIONS SCANNED

COMPLETE AMP/HR DISCHARGE TEST:

1. DISCHARGE: AMPS HOURS TEMP
 CHARGE: AMPS HOURS TEMP
 2. DISCHARGE: AMPS HOURS TEMP
 CHARGE: AMPS HOURS TEMP

(CAUTION: AVOID EXCESS GASSING AND TEMPERATURE)

DATE 2004-04-15

TEST COMPLETED BY Dr. R. Gw

REVISED 03-04-08 (W3)

Notes:

Storey Brook

03/15/04 13:14

Reference Value Entered:
---- Siemens (MHOS)Reference Temperature:
15 Degrees Celsius

----string statistics----

string Average : 1388s
string% of HI Jar: 083%

-----Jar Statistics-----

JAR	JAR--V	JAR--G	% REF
001	2.25V	1438s	-----
002	2.25V	1470s	-----
003	2.24V	1279s	-----
004	2.24V	1571s	-----
005	2.25V	1452s	-----
006	2.25V	1512s	-----
007	2.24V	0929s	-----
008	2.25V	1166s	-----
009	2.25V	1470s	-----
010	2.25V	1407s	-----
011	2.24V	1576s	-----
012	2.25V	1539s	-----
013	2.22V	1205s	-----
014	2.25V	1584s	-----
015	2.27V	1265s	-----
016	2.27V	1281s	-----
017	2.27V	1484s	-----
018	2.24V	1462s	-----
019	2.24V	1410s	-----
020	2.24V	1213s	-----
021	2.25V	1602s	-----
022	2.25V	1481s	-----
023	2.25V	1580s	-----
024	2.24V	1108s	-----
025	2.25V	1539s	-----
026	2.25V	0919s	-----
027	2.24V	1309s	-----
028	2.25V	1444s	-----
029	2.27V	1484s	-----
030	2.25V	1515s	-----
031	2.25V	1523s	-----
032	2.24V	1563s	-----
033	2.25V	1661s	-----
034	2.25V	1140s	-----
035	2.25V	1503s	-----
036	2.24V	1519s	-----
037	2.25V	1509s	-----
038	2.22V	1459s	-----
039	2.25V	1515s	-----
040	2.25V	1535s	-----
041	2.24V	1135s	-----
042	2.25V	1466s	-----
043	2.22V	0903s	-----
044	2.24V	1496s	-----
045	2.24V	1550s	-----
046	2.24V	1437s	-----
047	2.24V	1351s	-----
048	2.22V	1301s	-----
049	2.24V	0926s	-----
050	2.25V	1519s	-----
051	2.24V	1333s	-----
052	2.24V	0867s	-----
053	2.25V	1530s	-----
054	2.25V	1452s	-----
055	2.25V	1547s	-----
056	2.25V	1628s	-----
057	2.22V	1007s	-----

Lowest Jar:
052Highest Jar:
033

---Settle Information---

String	Warn%	(?):	70%
String	Fail%	(?):	60%
Jar	Warn%	(?):	70%
Jar	Fail%	(?):	60%
Volt	Var	Set:	02V

End of Log

2003 PM'S

FORM 11 - VRLA BATTERY CELL INSPECTION

Equipment ID: STB Batt. Bank Station: STB Type No.: 90A-09
 Manufacturer: Absolyte 119 Date: 2003/04/02 Readings By: BP-GN

Room Temperature: 18 °C
 Average Battery Temperature: 18 °C

Voltage of Battery Disconnected: — V
 Pilot Cell Voltage: 2.20 V
 Pilot Cell No.: 48

	(CHARGER #1)	(CHARGER #2)
Voltage Output of Charger (V):	<u>130</u> Found <u>130</u> Left	Found <u>—</u> Left <u>—</u>
DC Current Output of Charger (A):	<u>5</u> Found <u>5</u> Left	Found <u>—</u> Left <u>—</u>
AC Current Output of Charger (A):	<u>—</u> Found <u>—</u> Left	Found <u>—</u> Left <u>—</u>

Terminal DC Voltage to Ground (V): — Positive — Negative

CHECKS COMPLETED:

- ☒ Check the accuracy of the charger voltage meter against another meter.
 - ☒ Check each cell for cracks, corrosion on terminals, leaks, swelling.
 - ☒ Inspect the battery rack for signs of corrosion & battery leakage.
 - ☒ Clean, lubricate cell connections if required.
 - ☐ Conduct individual cell temperature (form V).
 - ☒ Check the operation of ventilation fans and thermostat.
 - ☒ Check the operation of emergency lighting.
 - ☒ Fire extinguisher (Class C) available & inspected.
 - ☒ Evaluate and attach MIDTRONICS Battery Diagnosis Test report.
- (Tests to be completed on Batteries (Jars) and intercell connections (Straps).

ONLY IF INSTRUCTED:

- ☐ Conduct resistance test of cell-cell connections (form VI).
- ☐ Collect cell internal ohmic values (form III).

**Notes:

125V Bank: Float - 129V
 48V Bank: Float - 51.6V

Alarm checked OK

microCELLTRON DataLogger

Notes: STBGN/BP2009-04-02

04/02/03 11:41

Reference Value Entered:
1530 Siemens (MH05)Reference Temperature:
18 Degrees Celsius

----string statistics----

String Average : 2788s
String% of Ref. : 190%
String% of HI Jar: 085%

-----Jar Statistics-----

JAR	JAR--V	JAR--G	+STRAP
001	2.25V	2970S	2806S
002	2.24V	3000S	2692S
003	2.25V	2642S	2512S
004	2.24V	2932S	2804S
005	2.27V	2968S	2830S
006	2.24V	3114S	2896S
007	2.25V	2228S	2118S
008	2.24V	2568S	2376S
009	2.25V	2924S	2802S
010	2.25V	2862S	2740S
011	2.25V	3134S	3000S
012	2.25V	3014S	2846S
013	2.24V	2666S	2524S
014	2.25V	3134S	2882S
015	2.27V	2798S	2648S
016	2.27V	2738S	2602S
017	2.27V	2948S	2806S
018	2.25V	3014S	2842S
019	2.24V	2824S	2624S
020	2.25V	2482S	2258S
021	2.25V	3152S	2954S
022	2.25V	2976S	2878S
023	2.25V	3168S	2992S
024	2.24V	2458S	2366S
025	2.25V	3070S	2840S
026	2.25V	2234S	2148S
027	2.24V	2692S	2562S
028	2.25V	2912S	2726S
029	2.27V	2976S	2806S
030	2.25V	3006S	2840S
031	2.25V	3046S	2848S
032	2.24V	2948S	2820S
033	2.25V	3264S	3086S
034	2.22V	2514S	2444S
035	2.24V	3046S	2862S
036	2.22V	2910S	2984S
037	2.25V	2774S	2836S
038	2.22V	2660S	3046S
039	2.25V	2882S	2992S
040	2.24V	2860S	2636S
041	2.22V	2518S	3014S
042	2.25V	2874S	2100S
043	2.22V	1954S	3000S
044	2.24V	2846S	3052S
045	2.24V	2982S	2924S
046	2.24V	2754S	2940S
047	2.24V	2734S	2496S
048	2.20V	2308S	2024S
049	2.24V	1946S	3046S
050	2.27V	2868S	2792S
051	2.24V	2672S	2112S
052	2.24V	2030S	3094S
053	2.25V	2924S	3100S
054	2.25V	2904S	3078S
055	2.24V	2896S	3322S
056	2.25V	3100S	2186S
057	2.24V	2098S	0000S

Best Jar: 043 133%

Highest Jar: 033 223%

---Setup Information---

String Warn% (?): 70%

2002 PM'S

LOCATION Stony BrookMANUF. Absolyte GNBCAT. NO. 90A-09TYPE ELECTROLYTE Dry (Gel)

CHARGING CONDITION FOUND:

FLOCAT ☒EQUALIZE ☐VOLTS 128.8AMPS 5

PILOT CELL:

NO. 48 (FOUND)SP. GRAVITY N/AVOLTS 2.196TEMP N/A DEG CWATER ADDED: YES ☐ NO ☒EQUALIZE CHARGE: VOLTS ☐AMP ☐ HR. ☐BANK CLEANED ☒CLEAN AND LUBRICATE CONNECTIONS ☒ALARM TESTED ☒NEW PILOT CELL: NO. ☐SP. GRAVITY ☐CHECK ACCURACY OF CHARGER METERS: ☒

RECORD ALL CELLS:

SP GR	VOLT	TEMP	SP GR	VOLT	TEMP	SP GR	VOLT	TEMP
1.	2.261		21.	2.264		41.	2.247	
2.	2.258		22.	2.264		42.	2.272	
3.	2.260		23.	2.266		43.	2.244	
4.	2.228		24.	2.261		44.	2.259	
5.	2.265		25.	2.270		45.	2.259	
6.	2.282		26.	2.265		46.	2.256	
7.	2.258		27.	2.261		47.	2.254	
8.	2.263		28.	2.263		48.	2.196	
9.	2.271		29.	2.274		49.	2.257	
10.	2.266		30.	2.263		50.	2.251	
11.	2.260		31.	2.265		51.	2.258	
12.	2.266		32.	2.260		52.	2.255	
13.	2.246		33.	2.266		53.	2.265	
14.	2.266		34.	2.264		54.	2.262	
15.	2.279		35.	2.261		55.	2.255	
16.	2.275		36.	2.256		56.	2.264	
17.	2.286		37.	2.262		57.	2.251	
18.	2.263		38.	2.245		58.		
19.	2.258		39.	2.264		59.		
20.	2.262		40.	2.260		60.		

ONLY IF INSTRUCTED:

TERMINATIONS SCANNED _____

COMPLETE AMP/HR DISCHARGE TEST:

1. DISCHARGE: AMPS _____ HOURS _____ TEMP _____
 CHARGE: AMPS _____ HOURS _____ TEMP _____
 2. DISCHARGE: AMPS _____ HOURS _____ TEMP _____
 CHARGE: AMPS _____ HOURS _____ TEMP _____

(CAUTION: AVOID EXCESS GASSING AND TEMPERATURE)

DATE 2002/02/06TEST COMPLETED BY BP-MG

REVISED 98-04-08 (WV)

LOCATION Stony BrookMANUF. Absolyte GNBCAT. NO. 90A-09TYPE ELECTROLYTE Dry

CHARGING CONDITION FOUND:

FLOAT

EQUALIZE ☒VOLTS 133.7AMPS 8

PILOT CELL:

NO. 48 (FOUND)SP. GRAVITY N/AVOLTS 2.194TEMP N/A

DEG C

WATER ADDED: YES

NO ☒EQUALIZE CHARGE: VOLTS 133.7AMP 8HR. 4BANK CLEANED ☒CLEAN AND LUBRICATE CONNECTIONS ☒ALARM TESTED ☒

NEW PILOT CELL: NO

SP. GRAVITY

CHECK ACCURACY OF CHARGER METERS: ☒

Readings After 4 Hrs of Equalize

RECORD ALL CELLS:

SP GR	VOLT	TEMP	SP GR	VOLT	TEMP	SP GR	VOLT	TEMP
1.	2.265		21.	2.264		41.	2.247	
2.	2.262		22.	2.266		42.	2.271	
3.	2.263		23.	2.269		43.	2.243	
4.	2.233		24.	2.264		44.	2.258	
5.	2.271		25.	2.273		45.	2.258	
6.	2.286		26.	2.268		46.	2.256	
7.	2.262		27.	2.264		47.	2.254	
8.	2.266		28.	2.266		48.	2.194	
9.	2.273		29.	2.276		49.	2.257	
10.	2.268		30.	2.265		50.	2.281	
11.	2.263		31.	2.265		51.	2.259	
12.	2.267		32.	2.260		52.	2.258	
13.	2.248		33.	2.267		53.	2.267	
14.	2.267		34.	2.265		54.	2.263	
15.	2.281		35.	2.261		55.	2.256	
16.	2.277		36.	2.255		56.	2.265	
17.	2.279		37.	2.261		57.	2.252	
18.	2.265		38.	2.245		58.		
19.	2.260		39.	2.264		59.		
20.	2.264		40.	2.261		60.		

ONLY IF INSTRUCTED:

TERMINATIONS SCANNED

COMPLETE AMP/HR DISCHARGE TEST:

1. DISCHARGE: AMPS _____ HOURS _____ TEMP _____
 CHARGE: AMPS _____ HOURS _____ TEMP _____
 2. DISCHARGE: AMPS _____ HOURS _____ TEMP _____
 CHARGE: AMPS _____ HOURS _____ TEMP _____

(CAUTION: AVOID EXCESS GASSING AND TEMPERATURE)

DATE 2002/02/07TEST COMPLETED BY BP-MG

REVISED 1980-04-08 (WTS)

After Equalize 128.9 VDS
 5 Amps

2001 PM'S

BATTERY BANK

BOOKS

Stomp Book

MANUF.

GNB Industrial Battery Co.

NO.

90A-09

TYPE ELECTROLYTE

Dry Fit Sulfuric Acid

ING CONDITION FOUND:

AT

✓

EQUALIZE

VOLTS 128.89

AMPS

6

CELL:

48

(FOUND)

SP. GRAVITY

N/A

VOLTS

2.17

TEMP

DEG C

ADDED: YES

NO

✓

EQUALIZE CHARGE: VOLTS

AMP

HR.

CLEANED

yes

CLEAN AND LUBRICATE CONNECTIONS

✓

ALARM TESTED

OK

PILOT CELL: NO.

48

SP. GRAVITY

ACCURACY OF CHARGER METERS:

✓ OK

RECORD ALL CELLS:

GR	VOLT	TEMP	SP GR	VOLT	TEMP	SP GR	VOLT	TEMP
	2.27		21.	2.25		41.	2.24	
	2.27		22.	2.28		42.	2.25	
	2.27		23.	2.25		43.	2.26	
	2.23		24.	2.27		44.	2.25	
	2.25		25.	2.26		45.	2.27	
	2.26		26.	2.25		46.	2.27	
	2.25		27.	2.27		47.	2.24	
	2.27		28.	2.26		48.	2.17	
	2.27		29.	2.28		49.	2.27	
	2.25		30.	2.25		50.	2.28	
	2.26		31.	2.27		51.	2.25	
	2.27		32.	2.27		52.	2.25	
	2.24		33.	2.25		53.	2.25	
	2.27		34.	2.28		54.	2.27	
	2.29		35.	2.25		55.	2.27	
	2.26		36.	2.27		56.	2.24	
	2.29		37.	2.27		57.	2.27	
	2.28		38.	2.27		59.		
	2.25		39.	2.25		59.		
	2.27		40.	2.27		60.		

INSTRUCTED:

CTIONS SCANNED

E AMP/HR DISCHARGE TEST:

CHARGE:	AMPS	HOURS	TEMP
ARGE:	AMPS	HOURS	TEMP
CHARGE:	AMPS	HOURS	TEMP
ARGE:	AMPS	HOURS	TEMP

(CAUTION: AVOID EXCESS GASSING AND TEMPERATURE)

00-01-31

TEST COMPLETED BY C. R. J. - W. F.

ED-98-04-08 (WV)

Western Avalon Battery PM'S

2006 PM'S

FORM II - FLOODED (VENTED) BATTERY CELL INSPECTION

Equipment ID: 129V Bank Station: WAV Serial No.: _____
 Manufacturer: Exide G Date: 2006/02/17 Readings By: Rob G.H.

Room Temperature: _____ °C
 Average Battery Temperature: _____ °C

Voltage of Battery Disconnected: N/A V
 Pilot Cell Voltage: 2.15 V
 Pilot Cell No.: #1 S.P. 1.210

(CHARGER #1)

(CHARGER #2)

Voltage Output of Charger (V):	<u>129</u> Found	<u>129</u> Left	Found _____	Left _____
DC Current Output of Charger (A):	<u>Temp</u> Found	<u>Temp</u> Left	Found _____	Left _____
AC Current Output of Charger (A):	<u>N/A</u> Found	<u>N/A</u> Left	Found _____	Left _____

Terminal DC Voltage to Ground (V): 65.8 Positive 62.2 Negative

CHECKS COMPLETED:

- ☒ Check the accuracy of the charger voltage meter against another meter.
 - ☒ Check each cell for cracks, corrosion on terminals, leaks.
 - ☒ Inspect the battery rack for signs of corrosion & battery leakage.
 - ☒ Clean and lubricate all battery and cell connections as needed.
 - ☒ Collect specific gravity, electrolyte level & temperature of cells (form IV).
 - ☒ Check the operation of ventilation fans and thermostat.
 - ☒ Check the operation of emergency lighting.
 - ☒ Fire extinguisher (Class C) available & inspected.
 - ☒ Evaluate and attach MIDTRONICS Battery Diagnosis Test report.
- (Tests to be completed on all Batteries and Cell Interconnections (Straps).

ONLY IF INSTRUCTED:

- N/A Apply an equalizing charge to battery.
- N/A Conduct resistance test of cell-cell connections (form VI).

**Notes:

125V Bank: Float - 129V Equalize - 139V
 48V Bank: Float - 51.6V Equalize - 56V

-Equalize charge is required after any water is added
 -Do not allow excessive gassing

FORM IV - Battery Cell Inspection

 System ID: 12902 Bank
 Manufacturer: Exide G

 Station: WAV
 Date: 2006/02/15

 Serial No.: _____
 Readings By: VBW, POG
Lead Ant; many

	Volt	Temp (°C)	Specific Gravity	Electrolyte Level			Volt.	Temp (°C)	Specific Gravity	Electrolyte Level	
				(low / normal / high)	Refilled					(low / normal / high)	Refilled
1			1.210			31			1.235		
2			1.235			32			1.235		
3			1.220			33			1.230		
4			1.225			34			1.250		
5			1.215			35			1.225		
6			1.225			36			1.250		
7			1.230			37			1.220		
8			1.240			38			1.240		
9			1.230			39			1.245		
10			1.235			40			1.235		
11			1.240			41			1.235		
12			1.225			42			1.235		
13			1.230			43			1.250		
14			1.230			44			1.235		
15			1.220			45			1.240		
16			1.230			46			1.220		
17			1.230			47			1.225		
18			1.240			48			1.225		
19			1.240			49			1.225		
20			1.230			50			1.235		
21			1.230			51			1.225		
22			1.235			52			1.220		
23			1.230			53			1.240		
24			1.235			54			1.230		
25			1.240			55			1.230		
26			1.230			56			1.235		
27			1.225			57			1.235		
28			1.235			58			1.230		
29			1.220			59			1.225		
30			1.225			60			1.230		

** Take temperature readings at the negative terminal of cells.

Take specific gravity readings before adding water to cells.

Add water to cells with low levels of electrolyte.



MicroSELETRON DataLogger

Notes: *WAD TS**Beats to a measurement*

15/02/06 13:44

Reference Value Entered:
#--- Siemens (MAGS)Reference Temperature:
14 Degrees Celsius

---String Statistics---

String Average : 941%
String% of HI Jar: 094%

---Jar Statistics---

JAR	JAR--V	JAR--G	+STRAP
001	6.45V	0975%	0969%
002	6.45V	0978%	0964%
003	6.43V	0966%	0921%
004	6.45V	0957%	0939%
005	6.43V	0948%	0913%
006	6.46V	0936%	0948%
007	6.45V	0938%	0918%
008	6.45V	0940%	0936%
009	6.46V	0963%	0945%
010	6.46V	0924%	0904%
011	6.49V	0936%	0933%
012	6.51V	0930%	0915%
013	6.45V	0897%	0873%
014	6.46V	0942%	0936%
015	6.49V	0927%	0913%
016	6.49V	0930%	0931%
017	6.48V	0945%	0930%
018	6.45V	0990%	0957%
019	6.40V	0870%	0867%
020	6.49V	0873%	0860%

Lowest Jar:
019Highest Jar:
016

---Setup Information---

String Warn% (??): 60%
 String Fail% (??): 50%
 Jar Warn% (??): 60%
 Jar Fail% (??): 30%
 Volts/Jar Selected: 06V

End of Data.

2004 PM'S

FORM II - FLOODED (VENTED) BATTERY CELL INSPECTION

Equipment ID: Battery Bank 129V Station: W.A.V. Serial No.: 51088
 Manufacturer: eydi C Date: 2004/01/21 Readings By: GH WJ.

Room Temperature: 12° °C
 Average Battery Temperature: 59°F °C

Voltage of Battery Disconnected: _____ V
 Pilot Cell Voltage: _____ V
 Pilot Cell No.: _____

	(CHARGER #1)		(CHARGER #2)	
Voltage Output of Charger (V):	<u>129</u> Found	<u>129</u> Left	Found _____	Left _____
DC Current Output of Charger (A):	<u>3</u> Found	<u>3</u> Left	Found _____	Left _____
AC Current Output of Charger (A):	<u>N/A</u> Found	<u>-</u> Left	Found _____	Left _____

Terminal DC Voltage to Ground (V): 64.7 Positive 64.2 Negative

CHECKS COMPLETED:

- ☒ Check the accuracy of the charger voltage meter against another meter.
 - ☒ Check each cell for cracks, corrosion on terminals, leaks.
 - ☒ Inspect the battery rack for signs of corrosion & battery leakage.
 - ☒ Clean and lubricate all battery and cell connections as needed.
 - ☒ Collect specific gravity, electrolyte level & temperature of cells (form IV).
 - ☒ Check the operation of ventilation fans and thermostat.
 - ☒ Check the operation of emergency lighting.
 - ☒ Fire extinguisher (Class C) available & inspected.
 - ☒ Evaluate and attach MIDTRONICS Battery Diagnosis Test report.
- (Tests to be completed on all Batteries and Cell Interconnections (Straps))

ONLY IF INSTRUCTED:

- ☐ Apply an equalizing charge to battery.
- ☐ Conduct resistance test of cell-cell connections (form VI).

*Notes:

125V Bank: Float - 129V Equalize - 139V
 48V Bank: Float - 51.6V Equalize - 56V

-Equalize charge is required after any water is added
 -Do not allow excessive gassing

FORM IV - Battery Cell Inspection

System ID: Buholvy Bank
 Manufacturer: Exide

Station: W.A.V
 Date: 2004/01/21

Serial No.: 57088
 Readings By: G.H. W.D.

	Volt	Temp (°F)	Specific Gravity	Electrolyte Level			Volt	Temp (°F)	Specific Gravity	Electrolyte Level	
				(low / normal / high)	Refilled					(low / normal / high)	Refilled
1	2.14	59	1204	low	yes	31	2.15	59	1214	low	yes
2	2.14	59	1218			32	2.14	59	1218		
3	2.14	59	1216			33	2.15	59	1216		
4	2.14	59	1210			34	2.14	59	1219		
5	2.14	60	1212			35	2.14	59	1220		
6	2.14	60	1215			36	2.14	59	1219		
7	2.14	60	1211			37	2.14	57	1208		
8	2.11	60	1208			38	2.14	57	1214		
9	2.12	60	1207			39	2.14	57	1213		
10	2.15	60	1211			40	2.14	57	1216		
11	2.14	60	1210			41	2.14	57	1214		
12	2.14	60	1210			42	2.15	57	1217		
13	2.14	60	1207			43	2.14	57	1213		
14	2.14	60	1205			44	2.14	57	1218		
15	2.14	60	1208			45	2.14	57	1214		
16	2.14	60	1213			46	2.14	59	1213		
17	2.14	59	1206			47	2.14	59	1216		
18	2.15	59	1211			48	2.15	59	1213		
19	2.14	59	1213			49	2.15	59	1211		
20	2.14	59	1210			50	2.14	59	1217		
21	2.14	59	1210			51	2.14	59	1219		
22	2.14	59	1219			52	2.12	59	1219		
23	2.14	59	1212			53	2.15	60	1217		
24	2.15	59	1215			54	2.15	60	1223		
25	2.14	59	1212			55	2.15	60	1215		
26	2.14	59	1210			56	2.15	60	1216		
27	2.14	59	1217			57	2.14	60	1219		
28	2.14	59	1218			58	2.15	60	1215		
29	2.14	59	1220			59	2.15	60	1226		
30	2.15	59	1209			60	2.14	60	1227		

** Take temperature readings at the negative terminal of cells.

Take specific gravity readings before adding water to cells.

Add water to cells with low levels of electrolyte.

Notes: *WAV*

Code: *C*

Unit: *W-D*

21/01/04 12:55

Reference Value Entered:

--- Siemens (MHOS)

Reference Temperature:

12 Degrees Celsius

---String Statistics---

String Average : 842s

String% of HI Jar: 884%

---Jar Statistics---

JAR JAR--V JAR--G % REF

001 2.14V 0900s

002 2.14V 0864s

003 2.14V 0911s

004 2.14V 0873s

005 2.14V 0892s

006 2.14V 0849s

007 2.14V 0874s

008 2.12V 0846s

009 2.15V 0857s

010 2.14V 0858s

011 2.14V 0872s

012 2.14V 0871s

013 2.14V 0866s

014 2.14V 0857s

015 2.14V 0829s

016 2.14V 0867s

017 2.15V 0883s

018 2.14V 0793s

019 2.14V 0872s

020 2.14V 0862s

021 2.14V 0863s

022 2.14V 0863s

023 2.14V 0852s

024 2.14V 0893s

025 2.14V 0877s

026 2.14V 0852s

027 2.14V 0841s

028 2.14V 0803s

029 2.15V 0844s

030 2.15V 0867s

031 2.14V 0893s

032 2.14V 0849s

033 2.14V 0849s

034 2.14V 0849s

035 2.14V 0852s

036 2.14V 0852s

037 2.14V 0825s

038 2.14V 0806s

039 2.14V 0815s

040 2.14V 0849s

041 2.14V 0863s

042 2.15V 0849s

043 2.14V 0860s

044 2.14V 0854s

045 2.14V 0843s

046 2.14V 0995s

047 2.14V 0886s

048 2.15V 0905s

049 2.15V 0849s

050 2.14V 0855s

051 2.14V 0852s

052 2.12V 0805s

053 2.15V 0806s

054 2.15V 0806s

055 2.15V 0814s

056 2.15V 0814s

057 2.14V 0785s

058 2.15V 0810s

059 2.15V 0821s

060 2.14V 0785s

Lowest Jar:

005

Highest Jar:

046

---Setup Information---

String Warn% (?) : 60%

String Fail% (?) : 50%

Jar Warn% (?) : 60%

Jar Fail% (?) : 50%

Volts/Jar Selected: 06V

End of Data

8-29-06; 11:38AM;Whitbourne

Notes: *WAV*

Code: *C*

Unit: *W-D*

21/01/04 13:37

Reference Value Entered:

--- Siemens (MHOS)

Reference Temperature:

12 Degrees Celsius

---String Statistics---

String Average : 923s

String% of HI Jar: 894%

---Jar Statistics---

JAR JAR--V JAR--G +STRAP

001 7.04V 0954s 0936s

002 6.99V 0945s 0936s

003 6.94V 0909s 0900s

004 6.94V 0948s 0912s

005 6.91V 0930s 0807s

006 7.07V 0930s 0915s

007 7.04V 0900s 0888s

008 7.05V 0933s 0915s

009 7.04V 0942s 0930s

010 7.08V 0912s 0780s

011 7.07V 0924s 0903s

012 7.10V 0918s 0894s

013 7.13V 0873s 0852s

014 7.08V 0948s 0909s

015 7.04V 0950s 0813s

016 7.07V 0978s 0957s

017 7.04V 0936s 0918s

018 7.04V 0954s 0930s

019 7.05V 0858s 0843s

020 7.07V 0649s 0852s

Lowest Jar:

020

Highest Jar:

016

---Setup Information---

String Warn% (?) : 60%

String Fail% (?) : 50%

Jar Warn% (?) : 60%

Jar Fail% (?) : 50%

Volts/Jar Selected: 06V

End of Data

2003 PM'S

FORM II - FLOODED (VENTED) BATTERY CELL INSPECTION

Equipment ID: 129VDCStation: WAVSerial No.: 93605Manufacturer: EXIDEDate: 2003/02/13Readings By: WB/GBRoom Temperature: 18 °C

Average Battery Temperature: _____ °C

Voltage of Battery Disconnected: _____ V

Pilot Cell Voltage: 2.17 VPilot Cell No.: 15Voltage Output of Charger (V): 131 Found 131 LeftDC Current Output of Charger (A): 7 Found 7 Left

AC Current Output of Charger (A): _____ Found _____ Left

Terminal DC Voltage to Ground (V): 65.56 Positive 65.02 Negative

HECKS COMPLETED:

- ☒ Check the accuracy of the charger voltage meter against another meter.
- ☒ Check each cell for cracks, corrosion on terminals, leaks.
- ☒ Inspect the battery rack for signs of corrosion & battery leakage.
- ☒ Clean, lubricate, & tighten all battery and cell connections as needed.
- ☒ Collect specific gravity, electrolyte level, temperature & voltage readings of cells (form IV).
- ☒ Check the operation of ventilation fans and thermostat.
- ☒ Check the operation of emergency lighting.
- ☒ Fire extinguisher (Class C) available & inspected.

ONLY IF INSTRUCTED:

- ☐ Apply an equalizing charge to battery.
- ☐ Complete capacity (discharge) test of battery.
- ☐ Conduct resistance test of cell-cell connections (form VI).

Equalizing Charge			
Why given:			
Voltage:	Amps:	Duration:	hrs
Capacity Test			
Initial		Final (Battery Voltage at Specified Minimum Value)	
Current:	Amps	Current:	Amps
Voltage:	Volts.	Voltage:	Volts.
Battery Temp:	°C	Battery Temp:	°C
Discharge Time		hrs	
Complete form V after discharge			

*Notes:

125V Bank: Float - 129V Equalize - 139V
 48V Bank: Float - 51.6V Equalize - 56V

-Equalize charge is required after any water is added
 -Do not allow excessive gassing

FORM IV - Battery Cell Inspection

System ID: 129VDC
 Manufacturer: EXIDE

Station: WAV
 Date: 2003/02/13

Serial No.: 93605
 Readings By: NO/GB

	Volt.	Temp (°C)	Specific Gravity	Electrolyte Level			Volt.	Temp (°C)	Specific Gravity	Electrolyte Level	
				(low / normal / high)	Refilled					(low / normal / high)	Refilled
1	2.18		1.19	N		31	2.18		1.20	N	
2	2.18		1.19	N		32	2.18		1.19	N	
3	2.17		1.19	N		33	2.18		1.18	N	
4	2.18		1.19	N		34	2.18		1.19	N	
5	2.18		1.19	N		35	2.18		1.18	N	
6	2.18		1.18	N		36	2.18		1.20	N	
7	2.17		1.19	N		37	2.20		1.18	N	
8	2.18		1.18	N		38	2.20		1.18	N	
9	2.17		1.19	N		39	2.17		1.20	N	
10	2.17		1.19	N		40	2.19		1.19	N	
11	2.17		1.19	N		41	2.18		1.18	N	
12	2.18		1.20	N		42	2.18		1.19	N	
13	2.17		1.19	N		43	2.18		1.19	N	
14	2.17		1.18	N		44	2.18		1.19	N	
15	2.17		1.17	N		45	2.18		1.18	N	
16	2.18		1.21	N		46	2.20		1.20	N	
17	2.18		1.20	N		47	2.18		1.20	N	
18	2.18		1.19	N		48	2.18		1.18	N	
19	2.18		1.19	N		49	2.16		1.20	N	
20	2.18		1.19	N		50	2.19		1.20	N	
21	2.18		1.19	N		51	2.18		1.20	N	
22	2.18		1.19	N		52	2.18		1.20	N	
23	2.17		1.20	N		53	2.18		1.20	N	
24	2.18		1.20	N		54	2.18		1.20	N	
25	2.18		1.19	N		55	2.18		1.20	N	
26	2.18		1.19	N		56	2.20		1.19	N	
27	2.18		1.20	N		57	2.20		1.19	N	
28	2.18		1.20	N		58	2.18		1.19	N	
29	2.18		1.19	N		59	2.19		1.19	N	
30	2.18		1.19	N		60	2.19		1.20	N	

** Take temperature readings at the negative terminal of cells.

Take specific gravity readings before adding water to cells.

Add water to cells with low levels of electrolyte.

2001 PM'S

6407

193479

STATION BATTERY & CHARGERS - FORM II

97 MAR 05

Date 2001/02/24 Inspected By Rob. G. K. Station WAIL
 Mfg. Exide/c Cat. # 93605 Type Electrolyte Lead Antimony
 CHARGING COND. FOUND: Float ☒ Equalize ☐ Volts 132.5 Amps 7
 PILOT CELL: # 37 (Found) Sp. Gr. 1.250 Volts 2.21 Temp 70°F C
 WATER ADDED: Yes ☒ No ☐ EQUALIZE CHARGE: Volts Amp Hr.
 Bank Cleaned yes Terminations Scanned yes Alarm Tested ☒
 RECORD ALL CELLS: New Pilot Cell - # 52 Sp. Gr. 1.230

	SP.GR	VOLTS	TEMP.°C		SP.GR	VOLTS	TEMP.°C		SP.GR	VOLTS	TEMP.°C
1	1.260	2.2	22	21	1.250	2.21	22	41	1.260	2.21	22
2	1.250	2.2		22	1.250	2.21		42	1.250	2.2	
3	1.250	2.19		23	1.250	2.2		43	1.250	2.21	
4	1.260	2.2		24	1.275	2.21		44	1.260	2.21	
5	1.260	2.19		25	1.260	2.21		45	1.250	2.2	
6	1.240	2.19		26	1.260	2.21		46	1.250	2.21	
7	1.250	2.2		27	1.275	2.2		47	1.250	2.2	
8	1.275	2.2		28	1.275	2.21		48	1.275	2.21	
9	1.260	2.2		29	1.250	2.21		49	1.250	2.2	
10	1.250	2.2		30	1.250	2.21		50	1.250	2.2	
11	1.260	2.2		31	1.250	2.21		51	1.250	2.21	
12	1.260	2.19		32	1.260	2.22		52	1.230	2.2	
13	1.250	2.19		33	1.250	2.21		53	1.250	2.21	
14	1.260	2.19		34	1.275	2.21		54	1.260	2.21	
15	1.260	2.19		35	1.250	2.21		55	1.275	2.21	
16	1.260	2.21		36	1.240	2.21		56	1.275	2.22	
17	1.275	2.2		37	1.250	2.21		57	1.250	2.21	
18	1.275	2.2		38	1.250	2.22		58	1.250	2.2	
19	1.260	2.19		39	1.250	2.18		59	1.260	2.22	
20	1.260	2.21		40	1.250	2.21		60	1.250	2.22	

ONLY IF INSTRUCTED

** CAUTION: AVOID EXCESS GASSING AND TEMPERATURE **

Clean & Lubricate Connections ☒ Check Accuracy of Charger Meters ☒

Complete Amp/Hr. Discharge Test:

1. Discharge:	Amps	_____	Hrs.	_____	Temp	_____
Charge:	Amps	_____	Hrs.	_____	Temp	_____
2. Discharge:	Amps	_____	Hrs.	_____	Temp	_____
Charge:	Amps	_____	Hrs.	_____	Temp	_____

2000 PM'S

STATION BATTERY - FORM II

1984-04-15

W/D 140800

STATION WAV INSPECTED BY P. O'Grady DATE 00, 04, 11
 MFG. EXIDE C CAT. NO. 93605 TYPE ELECTROLYTE Lead Antimony
 CHARGING CONDITION FOUND: FLOAT ☒ EQUALIZE ☐ VOLTS 132.6 AMPS 7
 PILOT CELL: NO. 37 (FOUND) SP. GR. 1.200 VOLTS 2.18 TEMP. 20 °C
 WATER ADDED: YES ☐ NO ☐ EQUALIZE CHARGE: VOLTS ☐ AMP ☐ HR. ☐
 BANK CLEANED ☐ TERMINATIONS SCANNED ☐ ALARM TESTED OK
 RECORD ALL CELLS: NEW PILOT CELL: NO. ☐ SP. GR. ☐

SP. GR.	VOLTS	TEMP.	SP. GR.	VOLTS	TEMP.	SP. GR.	VOLTS	TEMP.
1. <u>1.225</u>	<u>2.21</u>	<u>20°C</u>	21. <u>1.215</u>	<u>2.20</u>	<u>20°C</u>	41. <u>1.225</u>	<u>2.21</u>	<u>20°C</u>
2. <u>1.225</u>	<u>2.20</u>		22. <u>1.225</u>	<u>2.21</u>		42. <u>1.225</u>	<u>2.21</u>	
3. <u>1.230</u>	<u>2.21</u>		23. <u>1.230</u>	<u>2.20</u>		43. <u>1.225</u>	<u>2.20</u>	
4. <u>1.225</u>	<u>2.20</u>		24. <u>1.230</u>	<u>2.21</u>		44. <u>1.225</u>	<u>2.20</u>	
5. <u>1.230</u>	<u>2.20</u>		25. <u>1.230</u>	<u>2.21</u>		45. <u>1.225</u>	<u>2.20</u>	
6. <u>1.230</u>	<u>2.20</u>		26. <u>1.210</u>	<u>2.21</u>		46. <u>1.220</u>	<u>2.21</u>	
7. <u>1.235</u>	<u>2.19</u>		27. <u>1.235</u>	<u>2.21</u>		47. <u>1.225</u>	<u>2.21</u>	
8. <u>1.235</u>	<u>2.20</u>		28. <u>1.235</u>	<u>2.21</u>		48. <u>1.225</u>	<u>2.21</u>	
9. <u>1.235</u>	<u>2.20</u>		29. <u>1.225</u>	<u>2.21</u>		49. <u>1.225</u>	<u>2.20</u>	
10. <u>1.230</u>	<u>2.19</u>		30. <u>1.225</u>	<u>2.22</u>		50. <u>1.210</u>	<u>2.21</u>	
11. <u>1.230</u>	<u>2.20</u>		31. <u>1.230</u>	<u>2.21</u>		51. <u>1.225</u>	<u>2.21</u>	
12. <u>1.225</u>	<u>2.20</u>		32. <u>1.230</u>	<u>2.21</u>		52. <u>1.210</u>	<u>2.20</u>	
13. <u>1.225</u>	<u>2.20</u>		33. <u>1.225</u>	<u>2.21</u>		53. <u>1.220</u>	<u>2.21</u>	
14. <u>1.225</u>	<u>2.20</u>		34. <u>1.225</u>	<u>2.21</u>		54. <u>1.225</u>	<u>2.21</u>	
15. <u>1.225</u>	<u>2.19</u>		35. <u>1.225</u>	<u>2.21</u>		55. <u>1.225</u>	<u>2.22</u>	
16. <u>1.225</u>	<u>2.21</u>		36. <u>1.230</u>	<u>2.21</u>		56. <u>1.225</u>	<u>2.22</u>	
17. <u>1.230</u>	<u>2.21</u>		37. <u>1.200</u>	<u>2.18</u>		57. <u>1.200</u>	<u>2.22</u>	
18. <u>1.225</u>	<u>2.21</u>		38. <u>1.230</u>	<u>2.22</u>		58. <u>1.205</u>	<u>2.20</u>	
19. <u>1.230</u>	<u>2.20</u>		39. <u>1.225</u>	<u>2.18</u>		59. <u>1.225</u>	<u>2.22</u>	
20. <u>1.210</u>	<u>2.20</u>		40. <u>1.225</u>	<u>2.21</u>		60. <u>1.225</u>	<u>2.22</u>	

ONLY IF INSTRUCTED:

Clean & Lubricate Connections ☐ Check Accuracy of Charger Meters ☐

Complete Amp/Hr. Discharge Test:

1. Discharge:	Amps		Hrs.		Temp.	
Charge:	Amps		Hrs.		Temp.	
2. Discharge:	Amps		Hrs.		Temp.	
Charge:	Amps		Hrs.		Temp.	

CAUTION: AVOID EXCESS GASSING AND TEMPERATURE.